

Membrane Free Protocol

FORM STABLE GRAFTS

INCISION

Perform a mid-crest incision with single release no more than 3mm into the muco-gingival area. (Reflect enough to fully visualise defect)



FLAP

Stretch the flap to assist closure. Soft brushes separate periosteal fibres in larger defects without cutting.



SITE PREPARATION

Carefully degranulate the site and ensure bone bleed. (We recommend a [Degranulation kit](#): [See here](#))



APPLICATION

Apply direct from the syringe. Compress using fingers or moist gauze to ensure full defect conformity.

(No pre-mixing or setting time with [GelOs](#) or [Powerbone D Putty](#))



CLOSURE

Close by mesial then distal then central with single releasing incision. (See below if closure is not possible)



PROTECTION

Tissue healing and graft protection is recommended with a 'sacrificial' bandage if closure cannot be achieved.

(e.g., [Collagen Sponge](#) or Compressed [Parasorb](#) as shown)



REVIEW

When performing simultaneous implant/graft placement use implant stability measurement and repeat at review to support an early loading protocol. (e.g., [Penguin RFA](#))



([Powerbone Putty](#) is radiopaque to assist x ray follow-up)

Ivy Cottage, Bridgnorth Road,
Stourton, DY7 5BH

M. +44 (0) 7800 990 131
E. ask@regen-store.co.uk
W. www.regen-store.co.uk

Company registration number 10714405

 **Regen**
Store

Incorporating Ariston Dental UK